

A THREE-DIMENSIONAL BEAM PROFILE MONITOR BASED ON RESIDUAL GAS IONIZATION¹

D. Shapira, T.A. Lewis

A three-dimensional beam profile monitor based on tracking the ionization of the residual gas molecules in the evacuated beam pipe is described. Tracking in position and time of the ions and electrons produced in the ionization enables simultaneous position sampling in three dimensions. Special features which make it possible to sample very low beam currents were employed.

¹ Abstract of published paper: Nucl. Instrum. Meth. A Phys. Res. A **400** (1997) 185-194.